

Translated English of Chinese Standard: GB/T16886.11-2021

www.ChineseStandard.net → Buy True-PDF → Auto-delivery.

Sales@ChineseStandard.net

GB

NATIONAL STANDARD OF THE
PEOPLE'S REPUBLIC OF CHINA

ICS 11.100.20

CCS C 30

GB/T 16886.11-2021 / ISO 10993-11:2017

Replacing GB/T 16886.11-2011

Biological evaluation of medical devices -

Part 11: Tests for systemic toxicity

医疗器械生物学评价 第 11 部分：全身毒性试验

(ISO 10993-11:2017, IDT)

Issued on: November 26, 2021

Implemented on: December 01, 2022

Issued by: State Administration for Market Regulation;

Standardization Administration of the PRC.

Table of Contents

Foreword	4
Introduction	6
1 Scope	9
2 Normative references	9
3 Terms and definitions	9
4 General considerations	11
4.1 General	11
4.2 Selection of animal species	12
4.3 Animal status	12
4.4 Animal care and husbandry	12
4.5 Size and number of groups	13
4.5.1 Size of groups	13
4.5.2 Number of groups	13
4.5.3 Treatment controls	14
4.6 Route of exposure	14
4.7 Sample preparation	14
4.8 Dosing	14
4.8.1 Test sample administration	14
4.8.2 Dosage volumes	15
4.8.3 Dosage frequency	15
4.9 Body weight and food/water consumption	16
4.10 Clinical observations	16
4.11 Clinical pathology	16
4.12 Anatomic pathology	17
4.13 Study designs	17
4.14 Quality of investigation	17
5 Acute systemic toxicity	18
5.1 General	18
5.2 Study design	18
5.2.1 Preparations	18
5.2.2 Experimental animals	18
5.2.3 Test conditions	19
5.2.4 Body weights	19
5.2.5 Clinical observations	19
5.2.6 Pathology	20
5.3 Evaluation criteria	21
5.3.1 General	21
5.3.2 Evaluation of results	21
5.4 Final report	22

6 Repeated exposure systemic toxicity (subacute, subchronic and chronic systemic toxicity)	23
6.1 General	23
6.2 Study design	24
6.2.1 Preparations	24
6.2.2 Experimental animals	24
6.2.3 Test conditions	25
6.2.4 Body weights	25
6.2.5 Clinical observations	25
6.2.6 Pathology	26
6.3 Evaluation criteria	27
6.3.1 General	27
6.3.2 Evaluation of results	27
6.4 Final report	28
Annex A (Informative) Routes of administration	29
Annex B (Informative) Dosage volumes	32
Annex C (Informative) Common clinical signs and observations	33
Annex D (Informative) Suggested haematology, clinical chemistry and urinalysis measurements	34
Annex E (Informative) Suggested organ list for histopathological evaluation	36
Annex F (Informative) Organ list for limited histopathology for medical devices subjected to systemic toxicity testing	39
Annex G (Informative) Information on material-mediated pyrogens	41
Annex H (Informative) Subchronic rat systemic toxicity - Dual routes of parenteral administration	43
Bibliography	45

Foreword

This document is drafted in accordance with GB/T 1.1-2020 "Directives for standardization - Part 1: Rules for the structure and drafting of standardizing documents".

The document is Part 11 of GB/T 16886 "Biological evaluation of medical devices". GB/T 16886 has issued the following parts:

- Part 1: Evaluation and testing within a risk management process;
- Part 2: Animal welfare requirements;
- Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity;
- Part 4: Selection of tests for interactions with blood;
- Part 5: Tests for *in vitro* cytotoxicity;
- Part 6: Tests for local effects after implantation;
- Part 7: Ethylene oxide sterilization residuals;
- Part 9: Framework for identification and quantification of potential degradation products;
- Part 10: Tests for irritation and skin sensitization;
- Part 11: Tests for systemic toxicity;
- Part 12: Sample preparation and reference materials;
- Part 13: Identification and quantification of degradation products from polymeric medical devices;
- Part 14: Identification and quantification of degradation products from ceramics;
- Part 15: Identification and quantification of degradation products from metals and alloys;
- Part 16: Toxicokinetic study design for degradation products and leachables;
- Part 17: Establishment of allowable limits for leachable substances;
- Part 18: Chemical characterization of materials;
- Part 19: Physico-chemical, morphological and topographical characterization of materials;

- Part 20: Principles and methods for immunotoxicology testing of medical devices.

This document replaces GB/T 16886.11-2011 "Biological evaluation of medical devices - Part 11: Tests for systemic toxicity". Compared with GB/T 16886.11-2011, the main technical change is as follows:

- Modify the size of groups for chronic toxicity tests (see Table 1; Table 1 of the 2011 edition).

This document, using translation method, is identical to ISO 10993-11:2017 "Biological evaluation of medical devices - Part 11: Tests for systemic toxicity".

China's documents which have a consistent correspondence with the international documents normatively referenced in this document are as follows:

- GB/T 16886.1-2011 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009, IDT);
- GB/T 16886.2-2011 Biological evaluation of medical devices - Part 2: Animal welfare requirements (ISO 10993-2:2006, IDT).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. The issuing authority of this document shall not be held responsible for identifying any or all such patent rights.

This document was proposed by National Medical Products Administration.

This document shall be under the jurisdiction of National Technical Committee 248 on Biological Evaluation of Medical Devices of Standardization Administration of China (SAC/TC 248).

Drafting organizations of this document: Shandong Quality Inspection Center for Medical Devices, National Institutes for Food and Drug Control.

Main drafters of this document: Wang Xin, Sun Likui, Han Qianqian, Wang Han, Chu Xiangyu, Xu Jing.

The previous releases of this document and the documents it replaces are as follows:

- It was first issued in 1997 as GB/T 16886.11-1997 and first revised in 2011;
- This is the second revision.

Biological evaluation of medical devices -

Part 11: Tests for systemic toxicity

1 Scope

This document specifies requirements and gives guidance on procedures to be followed in the evaluation of the potential for medical device materials to cause adverse systemic reactions.

This document applies to systemic toxicity studies of medical devices or materials.

2 Normative references

The contents of the following documents, through normative references in this text, constitute indispensable provisions of this document. Among them, for dated references, only the edition corresponding to that date applies to this document. For undated references, the latest edition (including all amendments) applies to this document.

ISO 10993-1 Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process

ISO 10993-2 Biological evaluation of medical devices - Part 2: Animal welfare requirements

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10993-1 and the following apply.

3.1

Dose/dosage

Amount of test sample administered (e.g. mass, volume) expressed per unit of body weight or surface area.

3.2

Dose-effect

Relationship between the dosage and the magnitude of a defined biological effect either in an individual or in a population sample.

This is an excerpt of the PDF (Some pages are marked off intentionally)

Full-copy PDF can be purchased from 1 of 2 websites:

1. <https://www.ChineseStandard.us>

- SEARCH the standard ID, such as GB 4943.1-2022.
- Select your country (currency), for example: USA (USD); Germany (Euro).
- Full-copy of PDF (text-editable, true-PDF) can be downloaded in 9 seconds.
- Tax invoice can be downloaded in 9 seconds.
- Receiving emails in 9 seconds (with download links).

2. <https://www.ChineseStandard.net>

- SEARCH the standard ID, such as GB 4943.1-2022.
- Add to cart. Only accept USD (other currencies - <https://www.ChineseStandard.us>).
- Full-copy of PDF (text-editable, true-PDF) can be downloaded in 9 seconds.
- Receiving emails in 9 seconds (with PDFs attached, invoice and download links).

Translated by: Field Test Asia Pte. Ltd. (Incorporated & taxed in Singapore. Tax ID: 201302277C)

About Us (Goodwill, Policies, Fair Trading...): <https://www.chinesestandard.net/AboutUs.aspx>

Contact: Wayne Zheng, Sales@ChineseStandard.net

Linkin: <https://www.linkedin.com/in/waynezhengwenrui/>

----- The End -----